

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 21 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130119</u>	
Name <u>Tobin P Brennan</u>		ID# <u>4542</u>	Date <u>5-19-10</u>
A	Agency <u>Soldotna Police Department</u>		Phone # <u>(907) 262-4455</u>
Instrument Location <u>Soldotna Police Department (44510 Sterling Hwy, Soldotna Alaska 99669)</u>			
Alco S/N <u>X172985</u>		Target Value <u>.088</u>	High Pressure <u>1075</u>
B	Alco Test Values 1 st Alco <u>.086</u> 2 nd Alco <u>.087</u>		
Signature _____			
(OVER)			

SMB
5/24/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows. (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of June, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

DIAGNOSTIC CHECK

— BREATH ANALYSIS —

.008 ADJUSTED FOR 29.62 in	.007	22:13
ALCO TARGET	.007	22:13
BLANK TEST	.000	22:14
INTERNAL STANDARD	VERIFIED	22:14
ALCO TU 29.64 in	.006	22:14
BLANK TEST	.000	22:15
SUBJECT SAMPLE	.000	22:16
BLANK TEST	.000	22:16
ALCO TU 29.64 in	.007	22:17
BLANK TEST	.000	22:17

— DIAGNOSTIC CHECK —

CALIBRATION: OKAY

```
^#12'0)+,-.0123456789:;<=>@ABCDEFGHI  
JKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop  
qrstuvwxyz{|}~
```